ibbyoak.

INSTALLATION GUIDE



INSTALLATION GUIDE

TABLE OF CONTENTS

INTRODUCTION

Vibby DAK.

р 3	1 - Automatic fall detection
р 4	2 - The different modes
p 5	3 - Manuel set-up of wearing mode
	p 4

2 ACTIVE AND STORAGE MODE

Presentation of each mode	р6
Switch from one mode to another	p 7

3 TEST THE VIBBY OAK

Demo Mode

р8



Presentation of the application	р9
Connecting the Vibby OAK to the application	p 10-11
Use of the application	p 12

FACT SHEETS

3	1 - Automatic fall detection
4	2 - The different modes
5	3 - Manuel set-up of wearing mode
	4 - Verification of wearing mode
	5 - Manuel set-up of radio protocol
	6 - Battery replacement
	7 - Fitting the wrist strap /neck cord
	8 - Removing the wrist strap/neck cord
	9 - Problem solving
c	



PREAMBLE

Vibby OAK is a smart intentional alarm system that, in some cases, adds additional protection if the user is not able to manually raise the alarm.

Vibby Oak.

The Vibby OAK is designed to automatically detect heavy/ dangerous falls (sudden and quick loss of balance followed by a significant impact of the body with the floor leaving the user lying on the floor unable to recover to a standing position).

Fall detection technologies that underpin the automatic fall detection do not allow detection or interpretation of all fall situations.



INTRODUCTION



THE STEPS OF AUTOMATIC FALL DETECTION

VIBBY OAK: AN AUTOMATIC FALL DETECTOR[•]



*REMINDER:

The Vitalbase technology **does not allow the analysis all fall situations.**



INTRODUCTION

THE FUNCTIONS OF THE VIBBY OAK



Vibby Oak.



An evidenced **SOS button** is located in the middle of the Vibby OAK. By pressing this button the wearer **can himself raise an alarm**.





Multiple **sensors** combined with a new, innovative **algorithm** can **automatically raise** an alarm when **a heavy fall is detected**. (see practical fact sheet 1)



Vibby DAK.

ACTIVE MODE AND STORAGE MODE

PRESENTATION OF EACH MODE



What is the active mode?

When the Vibby OAK is in **active mode**, it means that it is **operational** and **ready to be worn**. The user benefits from the automatic fall detection and manual call functions.

How do I know if the Vibby OAK is in active mode?

Just push the button*, the Vibby OAK vibrates and the red LED illuminates.



* If the Vibby OAK is in neck worn mode, you have to press a little longer (half a second).



STORAGE MODE

What is the storage mode?

The Vibby OAK is shipped **by default** in storage mode. This means it is **deactivated** in order to **preserve battery life** during its storage and transport. If not used for a while, the Vibby OAK must be reset to storage mode (*see page 7*).

How do I know if the Vibby OAK is in storage mode?

By simply pressing the button, the Vibby OAK does not switch on and the device does not vibrate.

Quickly press the button twice. If the LED flashes (green or red) and vibrates, it means that the device is in storage mode.



IMPORTANT:

All subsequent set-up procedures need to be completed from storage mode.



ACTIVE MODE AND STORAGE MODE

SWITCH FROM ONE MODE TO ANOTHER

Familiarise yourself with all steps before starting the procedure.

SWITCH FROM ACTIVE MODE TO STORAGE MODE



Vibby Dak.

2

4

Put the Vibby OAK flat on a table. Make sure to keep it still throughout the 4 steps. Press the button until the red LED starts flashing and then release.

Before the red LED goes off, press the button quickly and successively 5 times. The Vibby OAK will vibrate and its LED flash once to confirm the switch to storage mode.

SWITCH FROM STORAGE MODE TO ACTIVE MODE



Press and hold the button until the red LED starts flashing.

2

Release the button while it is flashing. The Vibby OAK will now slightly vibrate to confirm



DEMO MODE

TEST THE VIBBY OAK

The demo mode is automatically activated for 5 minutes each time the Vibby OAK switches to active mode. The green LED flashes.

Vibby DAK.

The Vibby OAK integrates a demo mode because it is very difficult to simulate the real conditions of a fall. That is why Vitalbase integrated a demo mode.



The demo mode stops automatically after 5 minutes. The green LED goes off





PRESENTATION OF THE APPLICATION



Vibby Dak.

THE VIBBY OAK CAN BE EASILY SET UP BY USING A DEDICATED MOBILE APPLICATION:

> Wearing mode selection : wrist or pendant> Radio protocol selection

The application also displays essential information:

> Battery status

> Vibby OAK serial number for traceability

Radio ID for certain protocols



IMPORTANT:

TO CONNECT TO THE APPLICATION, YOU NEED A LOGIN AND A PASSWORD. THESE IDENTIFIERS WILL BE PROVIDED ON REQUEST BY CAREPHONE MANUFACTURER. PLEASE CONTACT:

support@vitalbase.fr



Vibby Dak.

No SIM ?

14:17

Welcome in the Vibby OAK"

Thanks to this application, access Inanks to this application, access the information and settings of the

SEARCHING FOR VIBBY OAK"

HOW TO CONNECT MY VIBBY ON "? >

0000000000

CONNECTING THE VIBBY OAK TO THE APPLICATION

Once identified*, you are redirected to the page « search for Vibby OAK » (see opposite).

All you need to do now to connect the Vibby OAK to the application is to follow the procedure on the next page.

***IMPORTANT:**

TO CONNECT TO THE APPLICATION, YOU NEED A LOGIN AND A PASSWORD. THESE IDENTIFIERS WILL BE PROVIDED ON REQUEST BY CAREPHONE MANUFACTURER. PLEASE CONTACT:

support@vitalbase.fr



CONNECTING THE VIBBY OAK TO THE APPLICATION

Familiarise yourself with all steps before starting the procedure.

You need to be identified by the application. The latter has to be in "search for Vibby OAK" mode:



Hold the button between the two series of flashes

VIBBY OAK" DETECTED

storage mode. (see pages 6 and 7)

Vibby Oak.

The Vibby OAK is connected to the mobile application.



USE OF THE APPLICATION











AUTOMATIC FALL DETECTION



SILENT ANALYSIS

After a fall, the Vibby OAK will analyse the person's movements for **20 seconds**. The alarm will be cancelled if and only if the person stands up and remains standing for at least **6 seconds. Even in the event of movements on the ground the alarm call will continue to be processed.**

ALERT

The Vibby OAK will vibrate and the red LED flashes intermittently for a period of **20 seconds**. From that moment on the alarm will be raised even if the person stands up. **The sending of the alarm can only be** cancelled by using the capacitive sensor.

ALARM CANCELLATION

As soon as the Vibby OAK starts to vibrate a capacitive sensor enables you to cancel the sending of a potentially false alarm. Just cover the Vibby OAK entirely with the palm of your hand for 2 seconds. Remove your hand as soon as vibrations stop. Otherwise the alarm will be sent.



THE DIFFERENT MODES



Set-up of **wearing mode** Set-up of **radio protocol**





STORAGE MODE

MANUAL SET-UP OF WEARING MODE







You have 1 minute to perform the next step (step 4). If the green LED goes off, go back to step 2.

Make sure that the Vibby OAK is in storage mode (see installation guide pages 6 and 7).

Press the button until the 2nd series of red flashes starts and then release immediately.

3 sec.

Press the button and hold it until the red LED lights up and then release immediately.





VERIFICATION OF WRIST / PENDANT MODE

HOW DO I KNOW IF THE VIBBY OAK IS IN WRIST MODE?





Make sure that the Vibby OAK is in storage mode (see installation guide pages 6 and 7).

Quickly press the button twice and then release.



The red LED flashes quickly.





The Vibby OAK vibrates.

HOW DO I KNOW IF THE VIBBY OAK IS IN PENDANT MODE?



Make sure that the Vibby OAK is in storage mode (see installation guide pages 6 and 7).



Quickly press the button twice and then release.



The green LED flashes quickly



The Vibby OAK vibrates.



MANUEL SET-UP OF PROTOCOLS



Make sure that the Vibby OAK is in storage mode (see installation guide pages 6 and 7).

Press the button until the 2nd series of red flashes starts and then release immediately.

You have 1 minute to perform the next step (step 4). If the green LED goes off, go back to step 2.



Press the button and hold it until the Vibby OAK will vibrate and then release immediately.



vitalizase LOT FOR A SAFER

Ex : protocol 2

BATTERY REPLACEMENT

Use the Vibby OAK battery **extracting tool** and replace the battery in 7 steps:



Take the extracting tool to remove the cover.



Make 1/4 of a turn with the tool.

Remove the cover.



Replace the used battery.





FITTING THE WRIST / PENDANT

The fitting of the wrist strap /neck cord of the Vibby OAK must be carried out by an installer in storage mode. The fitting is therefore deliberately difficult so that the user or his family cannot easily perform it themselves.



Place the Vibby OAK face down flat on the table. Squeeze the wrist strap firmly and exert a strong pressure (horizontal and then downward) until the wrist strap is correctly inserted (the retention features must be fully embedded in the enclosure).



REMOVING THE WRIST / PENDANT

The fitting of the wrist strap /neck cord of the Vibby OAK must be carried out by an installer in storage mode. The fitting is therefore deliberately difficult so that the user or his family cannot easily perform it themselves.





PROBLEM SOLVING

1. You are lost because the Vibby OAK is in an unknown status.

Probable cause:

Operating error during set-up.

Action(s):

- Wait for 5 min., make sure that the Vibby OAK is in storage mode and restart set-up from the beginning. - If you set up the Vibby OAK manually (wearing or protocol mode): try with the mobile application, it is much easier!

2. The application cannot be found in the Play Store/App Store or the installation fails.

Probable Cause:

The smartphone does not have the prerequisites. **Action:**

Check prerequisites and change smartphone if necessary.

Prerequisite Android:

Version 4.4+ and BLE (Bluetooth Low Energy) technology.

iOS (Apple): Iphone 4S and more recent models.

3. The application does not accept login/ password

Probable cause(s):

- The smartphone does not have internet.

- Your identifiers are incorrect.

Action(s):

- Activate WiFi or 3G/4G or just one of the two. Connections such as "freewifi" can cause problems. In this case, you have to switch to mobile data or activate geolocation. If you have a smartphone with IOS (Apple), then select "cellular data" in: settings > Vibby OAK.

- Check your identifiers and try to reconnect.

4. The application does not retain the login/ password.

Probable cause:

Error in the application.

Action(s):

Android: uninstall and reinstall the application.
iOS (Apple): go to « settings / Vibby OAK / » then select "keep the connection".

5. The application does not detect the Vibby OAK.

Probable cause:

Ill-defined smartphone authorisations or settings. Action(s):

- Make sure that the Vibby OAK > App has been activated (see page 9 of the installation guide).

- Make sure that the smartphone's Bluetooth connectivity is activated.

- Make sure that geolocation is authorised as well as any eventually requested authorisations. If you have an Android smartphone, go to: Settings > Applications > Vibby OAK > Authorisation.

6. The application detects the Vibby OAK but the screen remains frozen on "Vibby OAK detected"

Probable cause(s):

- Overloaded Bluetooth environment.

- The application dos not have the necessary rights.

- The customer account allowing you to connect is linked to a production code of the product. If the Vibby OAK is not assigned to the client's production code, it cannot be set up.

Action(s):

- Make sure that the Bluetooth environment is not overloaded: only connect one Vibby OAK at a time. There should only be one smartphone with an open application in your surroundings.

- Check your identifiers and try to reconnect.

- Contact Vitalbase to check client data.





The Vibby OAK is a smart intentional alarm system that, in some cases, adds additional protection if the user is not able to manually raise the alarm.

The Vibby OAK is designed to automatically detect heavy/dangerous falls (sudden and quick loss of balance followed by a significant impact of the body with the ground) leaving the user lying on the ground unable to recover to a standing position. Fall detection technologies that underpin the automatic fall detection do not allow the detection or interpretation of all fall situations.

Depending on where the Vibby OAK's is located on the user, the fall, even heavy, may not be detected for lack of sufficient shock or impact.

0

Zone Actopolis II 2 bis rue de Nully de Harcour 33610 CANEJAN - FRANCE



Tel : 05 57 35 63 70 **Fax :** 05 57 35 63 71



@vitalbasefrance

contact@vitalbase.fr